

October 13, 2023  
Nihon Kohden Corporation

## Response to Contec Wireless Product Vulnerabilities

Dear Customers,

Nihon Kohden has analyzed the vulnerabilities (CVE-2022-36158, CVE-2022-36159) which were included in the firmware of the Contec FLEXLAN FX3000/2000 series wireless products, and which were disclosed on September 1, 2022. As the conclusion of the analysis, it was determined that our products and systems are not affected by these vulnerabilities in their intended use or usage environment.

Nihon Kohden will continue monitoring the status of any vulnerabilities and share any information on the Nihon Kohden global website if we have any updates.

As a general guideline Nihon Kohden recommends all customers maintain good security practices and isolate NK equipment from other devices.

- 1) Please make sure that no unauthorized equipment is connected to the hospital network.
- 2) Please make sure that the equipment connected to the network is secured by separating the hospital network from external networks, such as by a firewall.

Please feel free contact us at the following URL with any questions and concerns you have.

[Contact Us | Nihon Kohden Global Site](#)