


**MK-3CL HEMATOLOGY CONTROL+CRP**

2.5mL


**ASSAY SHEET**

<b>LOT</b>	<b>MK3C232L</b>
	<b>2023-05-05</b>
<b>OPEN</b>	<b>14DAYS</b>


**MEK-1303 Barcode Entry**

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		MEK-1303	RANGE
<b>WBC</b>	10 <sup>3</sup> /μL	<b>2.01</b>	<b>± 0.60</b>
<b>RBC</b>	10 <sup>6</sup> /μL	<b>2.30</b>	<b>± 0.15</b>
<b>HGB</b>	g/dL	<b>5.91</b>	<b>± 0.40</b>
<b>HCT</b>	%	<b>19.8</b>	<b>± 2.0</b>
<b>MCV</b>	fL	<b>85.9</b>	<b>± 9.0</b>
<b>MCH</b>	pg	<b>25.7</b>	<b>± 2.4</b>
<b>MCHC</b>	g/dL	<b>29.8</b>	<b>± 3.0</b>
<b>PLT</b>	10 <sup>3</sup> /μL	<b>80.0</b>	<b>± 30.0</b>
<b>MPV</b>	fL	<b>9.3</b>	<b>± 3.0</b>
<b>RDW-CV</b>	%	<b>15.4</b>	<b>± 3.0</b>
<b>LY%</b>	%	<b>65.07</b>	<b>± 10.00</b>
<b>MO%</b>	%	<b>4.65</b>	<b>± 5.00</b>
<b>GR%</b>	%	<b>30.28</b>	<b>± 10.00</b>
<b>LY</b>	10 <sup>3</sup> /μL	<b>1.31</b>	<b>± 0.50</b>
<b>MO</b>	10 <sup>3</sup> /μL	<b>0.09</b>	<b>± 0.20</b>
<b>GR</b>	10 <sup>3</sup> /μL	<b>0.61</b>	<b>± 0.50</b>
<b>CRP</b>	Precision	<b>≤ 10.0%</b>	
		<b>(1.30-7.50 mg/L)</b>	

**Manual Registration**

Lot	<b>232</b>
Level	<b>Low</b>
Control	<b>MK-3C</b>


<https://www.nihonkohden.com/spi/index.html>

**NIHON KOHDEN CORPORATION**

1-31-4 NISHIOCHIAI, SHINJUKU-KU

TOKYO 161-8560 JAPAN



0694-901765A