

**MK-RE1**
**HEMATOLOGY CONTROL**  
**for reticulocyte**
**3.0mL**
**LEVEL 1**  
**ASSAY SHEET**

LOT No.	MKRE233-1
EXP. DATE	2023-05-05
OPEN VIAL STABILITY	14 DAYS

PARA-METER	UNIT	MEK -9200	RANGE
		ASSAY VALUES	
RET%	%	0.97	±1.00
RET	10 <sup>6</sup> /μL	0.0338	±0.0347
STAIN		RETICULONAC	

**MANUAL REGISTRATION**

When you register this control on manual, select the items and enter the values in the order below.  
 After that, enter the assay values and range on the right table or those calculated at your lab.

Control	MKRE					
Level	1					
LOT No.	233					
EXP. DATE	5	M	5	D	2023	Y

**QR CODE**

The accompanying handy bar-code reader to MEK-9200 can read the above QR Code and load assay values into its hematology analyzer.

The word "QR Code" is registered trademark of DENSO WAVE INCORPORATED in Japan and other countries.


**ATTENTIONS:**

- Store at 2-8°C
- When creating long-term control charts, a gradual downward trend may continue due to the characteristics of the control material.
- If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.
- Please note that the assay value ranges are reviewed lot by lot and are subject to change.



**NIHON KOHDEN CORPORATION**  
 1-31-4 NISHIOCHIAI, SHINJUKU-KU  
 TOKYO 161-8560 JAPAN



0694-901708N